

UCLA Principles of Neuroimaging

non-invasive brain stimulation with TMS and TDCS

transcranial magnetic stimulation (TMS)
(transcranial direct-current stimulation (TDCS))

Allan Wu, MD
Associate Director, TMS Lab, ALBMC
Dept of Neurology, UCLA
Feb 8, 2016

What are TMS and TDCS?

- ◆ Noninvasive neurostimulatory and neuromodulatory methods currently used in human subjects
- ◆ Modern era of use since 1985 for TMS, since 2001 for TDCS
- ◆ Brain mapping and clinical applications are growing
- ◆ Mechanisms are incompletely understood
- ◆ Animal & bioengineering models remain relatively uncommon

Clinical Neurophysiology 120 (2009) 2008–2039

Contents lists available at ScienceDirect

Clinical Neurophysiology

journal homepage: www.elsevier.com/locate/clinph

ELSEVIER

Guidelines

Safety, ethical considerations, and application guidelines for the use of transcranial magnetic stimulation in clinical practice and research[☆]

Simone Rossi^{a,*}, Mark Hallett^b, Paolo M. Rossini^{c,d}, Alvaro Pascual-Leone^e and The Safety of TMS Consensus Group¹

^aDipartimento di Neuroscienze, Sezione Neurologia, Università di Siena, Italy

^bHuman Motor Control Section, NINDS, NIH, Bethesda, USA

^cUniversità Campus BioMedico, Roma, Italy

^dCasa di Cura S. Raffaele, Cassino, Italy

^eBerenson-Alten Center for Noninvasive Brain Stimulation, Beth Israel Deaconess Medical Center and Harvard Medical School, Boston, USA

(transcranial) magnetic stimulation (TMS)






FDA approvals of rTMS for treatment of medication-refractory major depression




TDCS



Utz et al 2010 Neuropsychologia



What are TMS (magnetic stimulation) and TDCS (direct current stimulation)?





Wagner et al. (2007). *Annu Rev Biomed Eng*

TMS and TDCS


- ◆ **TMS**
- ◆ Faraday's principle of induction
- ◆ Brief stimulation period (~200 usec)
- ◆ Induces action potentials in nerve axons
- ◆ Repetitive and/or patterned stimulation can induce modulation
- ◆ **TDCS**
- ◆ Polarization across brain
- ◆ Long-duration 10-20 min DC stimulation
- ◆ Does not induce action potentials
- ◆ Modulates neural firing rates

Faraday's law of induction (TMS)


- A time-varying current (di/dt) in a wire loop will induce a magnetic field (B)
- The magnetic field will induce an electromotive force (ϵ) in an adjacent conductor



i




$\frac{di}{dt}$



ϵ

Walsh and Cowey 2000

TMS has intermediate temporal/spatial resolution but unique interference qualities




Nature Reviews | Neuroscience

Walsh and Cowey 2000

What does TMS stimulate?

- ◆ Coil geometry
- ◆ Coil placement
- ◆ Pulse waveform
- ◆ Coil orientation
- ◆ Pattern of stimulation
- ◆ Frequency TMS pulses
- ◆ Intensity of stimulation
- ◆ Duration of stimulation



What does TMS stimulate I? depends on coil

Circular coils









Figure-8 "focal" coils




Advances in TMS coil designs

Mini-coil





Tischler et al 2011, J Neurosci Meth

H-coil "deep" TMS



Fadini et al 2009


shield-plating



Gasca et al 2010, IEEE Eng Med Biol Soc


Boundary effects on magnetic stimulation

- ◆ Effects of magnetic stimulation depend on
- ◆ **Induced electromotive force (ϵ)**
- ◆ Charge separation on boundary limits




Boundary effects on magnetic stimulation

- ◆ Effects of magnetic stimulation depend on
- ◆ Induced electromotive force (ϵ)
- ◆ **Charge separation on boundary limits**




Boundary effects on magnetic stimulation

- ◆ TMS stimulation is parallel to scalp surface
- ◆ Lack of radial component to stimulation
- ◆ Inability to focus TMS effects in-depth




Deeper TMS coil designs





Huang et al (2009) Brain Stimul.

What does TMS stimulate II: tissue boundaries




- Induced currents depend on tissue inhomogeneities and boundaries
- sharper bends / shorter axons = lower thresholds




Wagner et al. (2004) IEEE Trans Biomed Eng

What does TMS stimulate II: axon boundaries




Nummenmaa et al. "Targeting of white matter tracts with transcranial magnetic stimulation." Brain Stimulation


What does TMS stimulate III?



- TMS preferentially produces trans-synaptic stimulation
- Compared to electrical stimulation, TMS responses are more variable and sensitive to both internal and external factors




TMS site




Sandrini et al 2011

Coil location: TMS hotspot and neuronavigation




TMS produces different waveforms




Kammer et al 2001


TMS effects depend on waveform & orientation



Mills et al 1992



Kammer et al 2001



Tings et al (2005), EBR


Common TMS study types

- Neurophysiology studies
 - Single-pulse TMS outcome measures (excitability)
 - Paired-pulse intra-cortical or cortico-cortical excitability
- Perturbation studies
 - Cortical perturbation (on-line, single-pulse or rTMS)
 - Cortical perturbation (off-line, “virtual lesion” or modulation)
- Modulatory effects of rTMS (e.g. plasticity effects)
 - After-effects of rTMS (neurophysiologic, behavioral, imaging)
 - Clinical trials of rTMS (single- or multisession)

Forms of TMS


- **Conventional**
- Single-pulse TMS
 - (1 pulse every 5-10 secs)
 - Paired-pulse TMS
 - Same vs different sites
- Repetitive TMS (rTMS)
 - Conventional rTMS
 - rTMS Low frequency rTMS (≤ 1 Hz)
 - High frequency rTMS (>5 Hz)
- **Non-conventional**
- Single-pulse TMS
 - State-dependent TMS
 - Paired-TMS or triggered-TMS
 - Paired-associative stimulation
- Repetitive TMS (rTMS)
 - Patterned rTMS
 - Theta-burst stimulation (rTMS 50 Hz triplets at 5 Hz)
 - Quadripulse Stimulation
 - Other

rTMS types



Rossi et al 2009


TMS protocols




Sandrini et al 2011

On-line vs off-line study designs

- “on-line” concurrent TMS/TDCS stimulation of ongoing process
 - Reliably (relatively) produces interpretable disruptive effects
 - Single pulses highly temporally specific
 - Can explain facilitative effects by models of competitive inhibition
 - Can yield measures of excitability over primary motor/visual cortex




- “off-line” rTMS/TDCS modulation method (?virtual lesion)
 - Avoids interference of on-line TMS with task
 - Temporo-spatial specificity poorer



Common TMS study types

- **Neurophysiology studies**
 - Single-pulse TMS outcome measures (excitability)
 - Paired-pulse intra-cortical or cortico-cortical excitability
- **Perturbation studies**
 - Cortical perturbation (on-line, single-pulse or rTMS)
 - Cortical perturbation (off-line, “virtual lesion” or modulation)
- **Modulatory effects of rTMS**
 - After-effects of rTMS (neurophysiologic, behavioral, imaging)
 - Clinical trials of rTMS (single- or multisession)

Neurophysiology TMS studies



Sandrini et al 2011

Cortical excitability

• Motor cortex excitability:

- Responsiveness of the motor cortex to stimulation
- Represents influences along the **cortico-spino-motor pathway**
- Attention, motor imagery, movement, learning, practice, action observation, emotions, afferent stimulation, drugs all can affect cortical excitability
- Outcome measures:
 - Motor threshold,
 - Motor evoked potential (MEP), Mapping motor (muscle) representation, Input-output curve,
 - Cortical silent period
 - Paired-pulse studies

• Visual cortex excitability:

- Responsiveness of the visual cortex to stimulation
- Outcome measures: Phosphene thresholds

Motor cortex excitability

Motor threshold (MT)


- Minimum stimulus intensity required to elicit a small motor response in a target muscle 50% of the time
- Can be assessed at rest (RMT) or active contraction (AMT)
- Enables comparable intensity of stimulation across subjects

Motor evoked potential (MEP)

- Motor responses in a target muscle evoked by TMS at a given suprathreshold intensity
- MEP size and latency can be quantified
- Most common measure of changes in cortical excitability


Relaxation:

Preinnervation: peak-to-peak < 50 μ V




Facilitation:

Preinnervation: 1-5% max. rms




Kaelin-Lang, J Neuro Methods 2000

TMS intensity and location in study of motor resonance during action observation




Fadiga et al 1995

Paired-pulse TMS can probe intracortical circuit excitability within motor cortex




Paired pulses assess inter-regional connectivity



Disorders with abnormal excitability



- Parkinson's disease
- Dystonia
- Stroke
- Epilepsy
- Depression
- Schizophrenia
- Essential tremor
- Amyotrophic lateral sclerosis
- Huntington's disease
- Tourette's syndrome
- Myelopathy
- Corticobasal gang degen
- Cerebellar degeneration
- Polyradiculoneuritis
- CNS demyelinating disease
- CNS tumors
- Restless leg syndrome
- Chronic fatigue syndrome
- Etc...

Perturbation TMS studies




Sandrini et al 2011

Single-pulse TMS over occipital lobe can disrupt visual perception




Amassian 1989 (Handbook of TMS 2002)

Visual cortex processing is necessary for Braille reading in the early blind subjects




Cohen et al 1997

Perturbation TMS studies




Sandrini et al 2011

Repetitive TMS



Sandrini et al 2011


Offline conventional rTMS modulation of cortical excitability



Touge et al (2001) Clin Neurophysiol


Maeda et al (2000) Exp Brain Res

Theta-burst stimulation



Huang et al (2005) Neuron


Back to TDCS...



Utz et al 2010 Neuropsychologia




Modeling TDCS



Wagner et al. (2007) Neuroimage & Wagner et al (2007). "" Annu Rev Biomed Eng

TDCS induces changes in motor excitability (MEP as outcome measure)



Nitsche & Paulus (2000) J Physiol

Effects of offline rTMS


- **Local effects**
 - **Increase (decrease) excitability to normalize abnormal excitability (or other physiologic measure)**
- **Distant effects**
 - Modulation of distant sites in a functional network (resting or state-related)
 - Decrease excitability to release inhibition in a distant area and achieve paradoxical facilitation (for example)
- **Cellular and molecular (neurotransmitter) effects**
 - Stimulate release (or modulate levels) of neurotransmitters
 - Modulation of signaling pathways and gene transcription

Effects of offline rTMS

- **Local effects**
 - Increase (decrease) excitability to normalize abnormal excitability (or other physiologic measure)
- **Distant effects**
 - **Modulation of distant sites in a functional network (resting or state-related)**
 - **Decrease excitability to release inhibition in a distant area and achieve paradoxical facilitation (for example)**
- **Cellular and molecular (neurotransmitter) effects**
 - **Stimulate release (or modulate levels) of neurotransmitters**
 - **Modulation of signaling pathways and gene transcription**


Virtual lesions and competitive inhibition

- Left hemispace neglect due to chronic right hemisphere lesions can be transiently improved with rTMS perturbations over left (unaffected) hemisphere



Oliveri et al 2001, Brighina et al 2003

Cellular and molecular mechanisms of TMS




- rTMS modulates
 - c-fos and c-jun expression
 - Possible BDNF mRNA expression
 - Dopamine, serotonin, vasopressin, others
- Effects may increase with daily rTMS

Arias-Carrion 2008

Common & other TMS study types


- Neurophysiology studies
 - Single-pulse TMS outcome measures (excitability)
 - Paired-pulse intra-cortical or cortico-cortical excitability
 - **State-dependent TMS and paired/triggered-TMS**
- Perturbation studies
 - Cortical perturbation (on-line, single-pulse or rTMS)
 - Cortical perturbation (off-line, “virtual lesion” or modulation)
- Modulatory effects of rTMS (**or other patterned TMS**)
 - After-effects of rTMS (neurophysiologic, behavioral, imaging)
 - Clinical trials of rTMS (single- or multisession)

State-dependency of TMS



Silvanto et al, TINS 2008


Types of neuromodulation to probe or shape plasticity



Shouval et al, Front Comput Neurosci 2010

Quartarone et al, TINS 2010


Paired associative stimulation (PAS)



- Electrical stimulation of median nerve is followed by a TMS pulse over sensorimotor cortex.
 - 90 pairs of stim-TMS are repeated every 20 sec
 - interstimulus interval 25 msec: facilitates selective MEP
 - linked to NMDA dependent LTP


Quartarone et al, Cur Op Neuro, 2008

Pair TMS with behavior (Hebbian learning)




Butefisch et al 2004, J Neurophys

Homeostatic plasticity (meta-plasticity) priming “state” before rTMS



Quartarone et al. 2006 TINS

Priming protocols and meta-plasticity



Siebner 2010, Clin Neurophysiol 121(4)

FDA approvals of rTMS for treatment of medication-refractory major depression



High-frequency rTMS for depression

- Randomized sham-controlled multicenter trial for rTMS
 - Left DLPFC rTMS 5 days per week, 4-6 weeks
 - 10 Hz rTMS (120% rMT), 4 sec on, then 26 sec rest
 - 143 active rTMS, 134 sham rTMS




Table 3. Adverse Events Occurring in the Active Treatment Group at a Rate of 5% or More and at Least Twice the Rate for Sham (with ME-Coded Preferred Terms Shown)


Body System Preferred term	Active TMS (n = 165) n (%)	Sham TMS (n = 158) n (%)
Eye disorders		
Eye pain	10 (6.1)	3 (1.9)
Gastrointestinal Disorders Toothache	12 (7.3)	1 (0.6)
General Disorders and Site Administration Conditions		
Application site discomfort	18 (10.9)	2 (1.3)
Application site pain	59 (35.8)	6 (3.8)
Facial pain	11 (6.7)	5 (3.2)
Musculoskeletal and connective tissue disorders		
Muscle twitching	34 (20.6)	5 (3.2)
Skin and subcutaneous tissue disorders		
Pain of skin	14 (8.5)	1 (0.6)

MedDRA, Medical Dictionary for Regulatory Activities.

O'Reardon et al (2007) Biol Psychiatry 62(11):

Can cortical modulation be directed to target specific symptoms?


- Motor circuit = motor symptoms
- Prefrontal circuit = mood symptoms



Obeso et al (2008) Mov Disord 23 Suppl 3: S548-559.


Magnetic Stimulation for the Treatment of Motor and Mood Symptoms of Parkinson's Disease (MASTER-PD trial)

- First prospective, double-blind, sham-controlled, parallel-group multicenter rTMS clinical trial in PD in North America
- Avoids medication side-effects and surgical risks
- Potential selectivity of effects (motor vs mood)
- Only multisession rTMS trial testing somatotopic effects of rTMS
- Realistic sham-rTMS conditions
- Rigorous safety and tolerability monitoring




Clinicaltrials.gov: NCT01080794

Magnetic Stimulation for the Treatment of Motor and Mood Symptoms of Parkinson's Disease (MASTER-PD)




- Patients with PD for >3 years, both motor (movement) symptoms and depression (with current or past treatment with an antidepressant)
- Outcome measures: UPDRS Part III (motor), HAM-D (mood/depression)
- Locations: Beth Israel Deaconess Medical Center (Harvard), UCLA, Toronto, Florida, Cleveland Clinic, Oregon, NYU
- Interim analysis: 450 patients screened, 71 patients enrolled, 58 with complete datasets

MASTER-PD - HAM-D depression score




Movement Disorders Congress 2015

MASTER-PD - UPDRS motor score



Movement Disorders Congress 2015

Sham rTMS conditions




Hamada et al, Movement Dis 2008

Magnetic Stimulation for the Treatment of Motor and Mood Symptoms of Parkinson's Disease (MASTER-PD)

- No significant differential effect on mood or motor function of real versus sham TMS.
- Multifocal M1 and/or DLPFC HF rTMS was no better than sham stimulation for motor or mood symptoms of PD.
- Sustained improvement of depression, regardless of stimulation status, points to universal benefit from study participation or from a perceived intervention.
- Transient improvement of UPDRS III indicates strong placebo response (perhaps related to salient electrical sham stimulation) preclude specific conclusions regarding multi-target rTMS efficacy.
- Better understanding of sham rTMS response in this particular population may help designing future efficacy studies.

Control conditions



Potential risks of rTMS

Known Risks

- ◆ Seizure induction
- ◆ Local pain and headache
- ◆ Hearing threshold shift
- ◆ Effects on cognition & mood
- ◆ Burns from scalp electrodes
- ◆ Metal in the head
- ◆ Other reported adverse events:
 - nausea, dental pain, fainting, pseudoseizures, tinnitus

Theoretical Risks

- ◆ Neurotoxicity
- ◆ Kindling
- ◆ Endocrine effects
- ◆ Social and psychological consequences of a seizure

Consensus statement on rTMS (Belmaker et al 2003)

- Those who administer rTMS should be trained as “first responders”
- rTMS should be performed in a medical setting with appropriate emergency facilities.
- Patients and research subjects should be continuously monitored
- participants should be informed of the risk of seizure and its possible medical and social consequences.
- dosage of rTMS should generally be limited by published safety guidelines (Wassermann et al 1998)

Current consensus risk assessment for TMS

- Absolute contraindication:
 - metallic hardware/implanted devices
- Increased / uncertain risks by TMS protocol
 - non-conventional rTMS including priming paradigms, long-lasting plasticity paradigms, multi-site TMS
 - Conventional high-frequency rTMS beyond safety parameters
- Increased / uncertain risk by subject
 - history of seizures, lesions of the brain, drugs that lower seizure threshold, sleep deprivation, alcoholism
- Uncertain risk due to other events
 - Pregnancy, severe or recent heart disease, implanted brain electrodes
- No risk category
 - None of above uncertain/increased risks
 - Single- or paired-pulse TMS
 - Conventional low- or high-frequency rTMS within safety parameters (intensity, frequency, train length, inter-train duration)

Comments about rTMS and neuromodulation

(Huang et al, Neuron, 2005)

- “The effectiveness of these paradigms raises ethical issues about the use of these methods in normal human subjects, who have nothing to gain from modulation of synaptic plasticity, in contrast to patients with particular neurological disorders.
- ..., so in addition to putting our proposed experimental methods before the ethics committee of our institution and gaining consent from subjects, we pursued the experiments in an incremental fashion starting with smaller intensities and lower frequencies of stimulation than those reported here.
- We found in all experiments that cortical excitability eventually returned to baseline, and no subject reported any side effects from experimentation.
- However, as methods for inducing plastic changes in human cortex become more powerful, such issues will require constant scrutiny and vigilance on the part of experimenters.”

TMS and TDCS

TMS	TDCS
Relatively expensive (\$50-100k)	Very inexpensive (\$250)
Safe but may induce seizures at high intensity or frequencies	Very safe
Causes scalp twitches, tapping on head, may be uncomfortable	Causes transient scalp tingling, very well tolerated
Moderate-sized effects	Very mild effects
Induces action potentials and interrupts ongoing neural activity	Does not induce action potentials, modulates firing rate of active neurons
Repetitive or patterned stimulation may induce modulation of neural excitability	Modulates neural excitability (facilitation or inhibition depending on polarity)

Future directions and applications of modeling TMS and TDCS effects

- ♦ TMS and TDCS are unique noninvasive methods of stimulating the human brain
- ♦ Most studies
 - TMS/TDCS as modulation/perturbation to interpret behavioral, neurophysiologic, clinical outcomes; some effects lasting.
- ♦ Gaps in knowledge
 - Mechanisms of effect (more realistic brain models, effects on networks/ connectivity, animal and tissue models)
 - Developing novel coils for focusing surface field (improved resolution) or deeper structures (greater effects)
 - Use as biomarker or surrogate marker for neuropsychiatric disorders of plasticity (not just function)
 - Predictive ability to predict response to potential invasive neurointerventions
 - Making TMS/TDCS as part of multimodal adjunctive treatment for neuropsychiatric disease